



Test Report: GSM60B18-P1J

60W AC-DC Reliable Green Medical Adaptor

■ DESIGN VERIFY TEST

Output Function Test
Input Function Test
Protection Function Test
Control Function Test
Component Stress Test

■ SAFETY & E.M.C. TEST

Safety Test
E.M.C. Test

■ RELIABILITY TEST

ENVIRONMENT TEST

■ DESIGN VERIFY TEST

OUTPUT FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	RIPPLE & NOISE	V1 : 120 mVp-p (Max)	I/P : 230VAC O/P : FULL LOAD Ta : 25°C	V1 : 73.2 mVp-p (Max)	P
2	OUTPUT VOLTAGE TOLERANCE	V1 : -3 %~ +3 % (Max)	I/P : 80 VAC / 264 VAC O/P : FULL/ MIN LOAD Ta : 25°C	V1 : -0.6297 %~ 0.6297 %	P
3	LINE REGULATION	V1 : -1 %~ +1 % (Max)	I/P : 100 VAC ~ 264 VAC O/P : FULL LOAD Ta : 25°C	V1 : 0 %~ 0 %	P
4	LOAD REGULATION	V1 : -3 %~ +3 % (Max)	I/P : 230 VAC O/P : FULL ~MIN LOAD Ta : 25°C	V1 : -0.6297 %~ 0.6297 %	P
5	SET UP TIME	230VAC : 1000 ms (Max) 115VAC : 1500 ms(Max)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	230VAC/ 459.508 ms 115VAC/ 1194684 ms	P
6	RISE TIME	230VAC : 30 ms (Max) 115VAC : 30 ms (Max)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	230VAC/ 7.839 ms 115VAC/ 8.961 ms	P
7	HOLD UP TIME	230VAC : 50 ms (TYP) 115VAC : 16 ms (TYP)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	230VAC/ 59.070 ms 115VAC/ 26.582 ms	P
8	OVER/UNDERSHOOT TEST	< ±5%	I/P : 230 VAC O/P : FULL LOAD Ta : 25°C	TEST : 0 %	P
9	DYNAMIC LOAD	V1 : 1800 mVp-p	I/P : 230 VAC (1).O/P : FULL /Min LOAD 90%DUTY/ 1KHZ (2).O/P : FULL /Min LOAD 90%DUTY/ 3KHZ (3).O/P : FULL /Min LOAD 90%DUTY/ 5KHZ (4).O/P : FULL /Min LOAD 50%DUTY/ 120HZ Ta : 25°C	(1) 326 mVp-p (2) 266 mVp-p (3) 264 mVp-p (4) 208 mVp-p	P

INPUT FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
----	-----------	---------------	----------------	--------	---------

1	INPUT VOLTAGE RANGE	80VAC~264 VAC	I/P : TESTING O/P : FULL LOAD Ta : 25°C	55.731 V~264V TEST : OK	P
			I/P : LOW-LINE-3V= 77 V HIGH-LINE+15%=300 V O/P : FULL/MIN LOAD ON : 30 Sec. OFF : 30 Sec 10MIN (AC POWER ON/OFF NO DAMAGE)		
2	INPUT FREQUENCY RANGE	47HZ ~63 HZ NO DAMAGE OSC	I/P : 80 VAC ~ 264 VAC O/P : FULL-MIN LOAD Ta : 25°C	TEST : OK	P
3	EFFICIENCY	89 % (TYP)	I/P : 230 VAC O/P : FULL LOAD Ta : 25°C	91.6 %	P
4	INPUT CURRENT	230V/ 1 A (TYP) 115V/ 1.4 A (TYP)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	I = 0.500 A/ 230 VAC I = 1.019 A/ 115 VAC	P
5	INRUSH CURRENT	230V/ 60 A (TYP) 115V/ 30 A (TYP) COLD START	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	I = 41.601 A/ 230 VAC I = 20.475 A/ 230 VAC	P
6	LEAKAGE CURRENT	< 50 uA / 264VAC	I/P : 264 VAC O/P : Min LOAD Ta : 25°C	FOR PATIENT 39.2 μA	P
7	NO LOAD CONSUMPTION PS-ON SHORT	< 0.1 W	I/P : 240VAC O/P : NO LOAD Ta : 25°C	< 0.0536 W	P

PROTECTION FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	OVER LOAD PROTECTION	105 % ~160 %	I/P : 230 VAC I/P : 115 VAC O/P : TESTING Ta : 25°C	149.8 %/ 230 VAC 142.7 %/ 115 VAC Protection type : Hiccup mode, recovers automatically after fault condition is removed	P
2	OVER VOLTAGE PROTECTION	CH1 : 18.9 V ~24.3 V	I/P : 230 VAC I/P : 115 VAC O/P : MIN LOAD Ta : 25°C	21.8 V/ 230 VAC 21.3 V/ 115 VAC Protection type : Shut down o/p voltage, re-power on to recover	P
3	OVER TEMPERATURE PROTECTION	SPEC : RTH2>70°C NO DAMAGE	I/P : 230 VAC O/P : FULL LOAD	O.T.P. Active Shut down o/p voltage, re-power on to recover	P



60W AC-DC Reliable Green Medical Adaptor

GSM60B series

4	SHORT PROTECTION	SHORT EVERY OUTPUT 1 HOUR NO DAMAGE	I/P : 264 VAC O/P : FULL LOAD Ta : 25°C	NO DAMAGE Hiccup Mode	P
---	------------------	--	---	--------------------------	---

CONTROL FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	ERP STEP2 COMPLIANT	LEVEL $V \geq 88\%$	I/P: 230 VAC/115VAC O/P:100/75/50/25/0% LOAD Ta:25°C	230V 89.856% 115V 88.835%	P

COMPONENT STRESS TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	Power Transistor (D to S) or (C to E) Peak Voltage	Q1 Rated : 700 V 10 A	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on (2) Output Short (3)Full load continue Ta : 25°C	(1) 604 V (2) 504 V (3) 584 V	P
2	Diode Peak Voltage	D100 Rated : 100 V 20 A	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on (2)Output Short (3)Full load continue Ta : 25°C	(1) 84.8 V (2) 65.6 V (3) 78.8 V	P
3	Input Capacitor Voltage	C 5 Rated : 120u /400V/105°C	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on /Off (2) Min load Turn on /Off (3)Full Load /Min load Change Ta : 25°C	(1) 376 V (2) 376 V (3) 376 V	P
4	Control IC Voltage Test	U 1 Rated : 28 V	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on /Off (2) Min load Turn on /Off (3)Full Load /Min load Change Ta : 25°C	(1) 17.1 V (2) 17.0 V (3) 15.1 V	P
5	CLAMP DIODE	D 1 Rated : 800 V 2 A	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on (2) Output Short (3)Full load continue Ta : 25°C	(3) 488 V (4) 426 V (3) 484 V	P

SAFETY & E.M.C. TEST

SAFETY TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	WITHSTAND VOLTAGE	I/P-O/P : 4 KVAC/min	I/P-O/P : 4.2KVAC/min	I/P-O/P : 1.756 mA	P



60W AC-DC Reliable Green Medical Adaptor

GSM60B series

			Ta : 25°C	NO DAMAGE	
2	ISOLATION RESISTANCE	I/P-O/P : 500VDC>100MΩ	I/P-O/P : 500 VDC Ta : 25°C/70% RH	I/P-O/P : 9999 MΩ NO DAMAGE	P

E.M.C TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	HARMONIC	BS EN/EN61000-3-2 CLASS A	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	PASS	P
2	CONDUCTION	BS EN/EN55011 (CISPR11), FCC PART 15 /CISPR22, CAN ICES-3(B)/NMB-3(B) CLASS B	I/P : 230 VAC (50HZ) O/P : FULL/50% LOAD Ta : 25°C	PASS Test by certified Lab	P
3	RADIATION	BS EN/EN55011 (CISPR11), FCC PART 15 /CISPR22, CAN ICES-3(B)/NMB-3(B) CLASS B	I/P : 230 VAC (50HZ) O/P : FULL LOAD Ta : 25°C	PASS Test by certified Lab	P
4	E.S.D	BS EN/EN61000-4-2 AIR : 15KV / Contact : 8KV	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	CRITERIA A	P
5	E.F.T	BS EN/EN61000-4-4 INPUT: 2KV	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	CRITERIA A	P
6	SURGE	BS EN/EN61000-4-5 L-N : 1KV	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	CRITERIA A	P
7	Test by certified Lab & Test Report Prepare				

RELIABILITY TEST

ENVIRONMENT TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
----	-----------	---------------	----------------	--------	---------

1	TEMPERATURE RISE TEST	<p>MODEL: GSM60B24-P1J</p> <p>1. ROOM AMBIENT BURN-IN : 1 HRS I/P : 230VAC O/P : FULL LOAD Ta=21.5°C</p> <p>2. HIGH AMBIENT BURN-IN : 2HRS I/P : 230VAC O/P : FULL LOAD Ta=47.1°C</p> <table border="1" data-bbox="662 481 1181 1041"> <thead> <tr> <th>NO</th> <th>Position</th> <th>ROOM AMBIENT Ta=21.5°C</th> <th>HIGH AMBIENT Ta=47.1 °C</th> </tr> </thead> <tbody> <tr><td>1</td><td>LF1</td><td>48.5°C</td><td>71.5°C</td></tr> <tr><td>2</td><td>LF2</td><td>58.5°C</td><td>80.4°C</td></tr> <tr><td>3</td><td>BD1</td><td>61.3°C</td><td>83.0°C</td></tr> <tr><td>4</td><td>C5</td><td>54.4°C</td><td>77.4°C</td></tr> <tr><td>5</td><td>D1</td><td>61.3°C</td><td>84.4°C</td></tr> <tr><td>6</td><td>D40</td><td>54.8°C</td><td>78.2°C</td></tr> <tr><td>7</td><td>RTH2</td><td>55.9°C</td><td>79.0°C</td></tr> <tr><td>8</td><td>T1</td><td>58.0°C</td><td>80.8°C</td></tr> <tr><td>9</td><td>C105</td><td>51.5°C</td><td>74.5°C</td></tr> <tr><td>10</td><td>D100</td><td>61.2°C</td><td>83.8°C</td></tr> <tr><td>11</td><td>C40</td><td>54.6°C</td><td>77.9°C</td></tr> <tr><td>12</td><td>Q1</td><td>56.5°C</td><td>79.7°C</td></tr> <tr><td>13</td><td>D42</td><td>50.3°C</td><td>73.2°C</td></tr> <tr><td>14</td><td>U1</td><td>46.9°C</td><td>70.6°C</td></tr> </tbody> </table>			NO	Position	ROOM AMBIENT Ta=21.5°C	HIGH AMBIENT Ta=47.1 °C	1	LF1	48.5°C	71.5°C	2	LF2	58.5°C	80.4°C	3	BD1	61.3°C	83.0°C	4	C5	54.4°C	77.4°C	5	D1	61.3°C	84.4°C	6	D40	54.8°C	78.2°C	7	RTH2	55.9°C	79.0°C	8	T1	58.0°C	80.8°C	9	C105	51.5°C	74.5°C	10	D100	61.2°C	83.8°C	11	C40	54.6°C	77.9°C	12	Q1	56.5°C	79.7°C	13	D42	50.3°C	73.2°C	14	U1	46.9°C	70.6°C	p
NO	Position	ROOM AMBIENT Ta=21.5°C	HIGH AMBIENT Ta=47.1 °C																																																														
1	LF1	48.5°C	71.5°C																																																														
2	LF2	58.5°C	80.4°C																																																														
3	BD1	61.3°C	83.0°C																																																														
4	C5	54.4°C	77.4°C																																																														
5	D1	61.3°C	84.4°C																																																														
6	D40	54.8°C	78.2°C																																																														
7	RTH2	55.9°C	79.0°C																																																														
8	T1	58.0°C	80.8°C																																																														
9	C105	51.5°C	74.5°C																																																														
10	D100	61.2°C	83.8°C																																																														
11	C40	54.6°C	77.9°C																																																														
12	Q1	56.5°C	79.7°C																																																														
13	D42	50.3°C	73.2°C																																																														
14	U1	46.9°C	70.6°C																																																														
2	OVER LOAD BURN-IN TEST	NO DAMAGE 1 HOUR (MIN)	I/P : 230 VAC O/P : 120% LOAD Ta : 25°C	TEST : OK	p																																																												
3	LOW TEMPERATURE TURN ON TEST	TURN ON AFTER 2 HOUR	I/P : 264VAC/100VAC O/P : 100 % LOAD Ta= -30°C	TEST : OK	p																																																												
4	HIGH HUMIDITY HIGH TEMPERATURE HIGH VOLTAGE TURN ON TEST	AFTER 12 HOURS IN CHAMBER ON CONTROL 40°C NO DAMAGE	I/P : 272 VAC O/P : FULL LOAD Ta=40°C HUMIDITY= 95 %R.H	TEST : OK	p																																																												
5	TEMPERATURE COEFFICIENT	±0.03%/°C (0-50°C)	I/P : 230 VAC O/P : FULL LOAD	± 0.005%/°C (0-50°C)	p																																																												
6	STORAGE TEMPERATURE TEST	<p>1. Thermal shock Temperature : -45°C~ +90°C</p> <p>2. Temperature change rate : 25°C / MIN</p> <p>3. Dwell time low and high temperature : 30 MIN/EACH</p> <p>4. Total test cycle : 5 CYCLE</p> <p>5. Input/Output condition : STATIC</p>			OK																																																												
7	THERMAL SHOCK TEST	<p>1. Thermal shock Temperature : -30°C~ +60°C</p> <p>2. Temperature change rate : 25°C / MIN</p> <p>3. Dwell time low and high temperature : 30 MIN/EACH</p> <p>4. Total test cycle : 10 CYCLE</p> <p>5. Input/Output condition : 230VAC/Full Load AC ON/OFF TEST turn on 58sec ; turn off 2sec</p>			OK																																																												



60W AC-DC Reliable Green Medical Adaptor

GSM60B series

8	VIBRATION TEST	1 Carton & 1 Set (1) Waveform : Sine Wave (2) Frequency : 10~500Hz (3) Sweep Time : 12min/sweep cycle (4) Acceleration : 2G (5) Test Time : 60min in each axis (X.Y.Z) (6) Ta : 25°C	TEST : OK	p
9	CAPACITOR LIFE CYCLE	SUPPOSE C105 IS THE MOST CRITICAL COMPONENT (1) I/P : 230VAC O/P : FULL LOAD Ta=25°C LIFE TIME (2) I/P : 230VAC O/P : FULL LOAD Ta=40°C LIFE TIME (3) I/P : 230VAC O/P : 75% LOAD Ta=40°C LIFE TIME (4) I/P : 230VAC O/P : 50% LOAD Ta=40°C LIFE TIME	(1) 265888HRS (2) 82425HRS (3) 115744HRS (4) 182881HRS	p
10	MTBF	3491.2K hrs min. Telcordia SR-332 (Bellcore) ; 694.3K hrs min. MIL-HDBK-217F (25°C)		p
11	DMTBF/Accelerated Life Test	Demonstration Mean Time Between Failure (Expected Life): Above 50,000 hours @ TA 50°C		p

SAMPLE	TEST RESULT	TESTER	APPROVAL
PRODUCT SAMPLE	PASS	XUJ	WANGDZ